

HEALTH ALERT NETWORK | HEALTH ADVISORY | Nov. 1, 2022

## Nontuberculous *Mycobacteria* Infections – Dental Water Source

The North Dakota Department of Health and Human Services (HHS) is providing this information from the Centers for Disease Control and Prevention (CDC) to dental and medical providers in North Dakota for awareness.

HHS encourages dental providers to review protocols and procedures to maintain safe water supplies and reduce the risk for infections following dental procedures.

We are asking providers to report post-procedure invasive infections due to nontuberculous *Mycobacteria*.

Reports can be made by calling 1-800-472-2180 or 701-328-2378.

Questions regarding reporting or infection prevention policies and procedures can be directed to the HHS using the same phone numbers.

### This is an official **CDC HEALTH ADVISORY**

## Outbreaks of Nontuberculous *Mycobacteria* Infections Highlight Importance of Maintaining and Monitoring Dental Waterlines

Distributed via the CDC Health Alert Network  
October 31, 2022, 1:00 PM ET  
CDCHAN-00478

### **Summary**

The Centers for Disease Control and Prevention (CDC) is issuing this Health Alert Network (HAN) Health Advisory to emphasize the importance of following existing recommendations for maintaining and monitoring dental waterlines. Multiple outbreaks of nontuberculous *Mycobacteria* (NTM) infections have occurred in children who received pulpotomies in pediatric dental clinics where the dental treatment water contained high levels of bacteria. CDC provides guidelines on infection control in dental settings which contain recommendations to treat dental unit waterlines and monitor water quality. Dental providers should be familiar with these recommendations on how to properly maintain and monitor their dental equipment to ensure that dental treatment water is safe for patient care.

### **Background**

While rare, there have been multiple documented cases of disease transmission from dental unit

waterlines (narrow-bore plastic tubing that carry water to the high-speed handpiece, air/water syringe, and ultrasonic scaler).<sup>1-5</sup> Dental units have unique characteristics that make them prone to biofilm formation.<sup>6</sup> Biofilms occur in dental unit waterlines due to the long, small-diameter tubing and low flow rates used in dentistry and the frequent periods of stagnation. As a result, high numbers of common waterborne bacteria can be found in untreated dental unit water systems. Disease-causing microorganisms found in untreated dental unit water can include *Legionella*, *Pseudomonas aeruginosa*, and nontuberculous *Mycobacteria* (NTM).

Dental providers and patients could be placed at risk of adverse health effects if dental unit water is not appropriately treated. In March 2022, CDC was notified of a new cluster of suspected NTM infections in children following dental procedures at a pediatric dental clinic. Investigation into this cluster is currently ongoing, and preliminary site visit data report that dental unit waterline testing results showed microbial counts much higher than the level recommended by CDC.

In 2016, an outbreak occurred at a pediatric dental clinic in Orange County, California, with 71 patients identified as having odontogenic NTM infections following pulpotomy procedures.<sup>1</sup> Municipal water stored in a pressurized bladder holding tank was used to fill the dental unit water bottles. The clinic was not using disinfectants on their dental unit waterlines or regularly monitoring water quality. All water samples tested from the dental units showed microbial counts higher than the level recommended by CDC and multiple species of NTM were identified in syringe water samples from five of the six treatment rooms. In 2015, 24 cases of odontogenic NTM infections were reported in children receiving pulpotomy treatment from a pediatric dental clinic in Georgia.<sup>2</sup> Investigators from the Georgia Department of Public Health found that municipal water was used during dental procedures, the clinic was not using a disinfectant in their dental unit waterlines, and the clinic was not regularly monitoring the water quality as [recommended by CDC](#). Microbial testing of the water samples taken from the dental units showed very high microbial counts of *Mycobacterium abscessus*. The *M. abscessus* isolates recovered from the water were found to be identical to eight isolates from tissue samples from seven of the patients, suggesting that water was the source of the infections.

The outbreaks in California and Georgia involved young children, with ages ranging from 4 to 8 years. Many of the children developed severe infections with clinical diagnoses such as cervical lymphadenitis and mandibular or maxillary osteomyelitis, and required hospitalization, treatments such as intravenous antibiotics, and surgical procedures. Complications from their infections included permanent tooth loss, hearing loss, facial nerve palsy, and incision fibrosis.

Because of the potential to form biofilm, CDC recommends that all dental unit waterlines be treated regularly with disinfectants to meet the Environmental Protection Agency (EPA) regulatory standards for drinking water (i.e.,  $\leq 500$  colony forming units (CFU)/mL of heterotrophic water bacteria).<sup>7</sup> There are many commercial products and devices available to disinfect and maintain dental unit waterlines. Dental unit water quality must also be monitored routinely as recommended by the equipment manufacturer to ensure that treatments are working effectively and that the water used in dental procedures meets safety standards.<sup>6</sup> Dental providers should consult with the dental equipment manufacturer for appropriate methods and equipment to both [maintain and monitor the quality of dental water](#).

Oral surgical procedures involve the incision, excision, or reflection of tissue that exposes the normally sterile areas of the oral cavity. Examples include biopsy, periodontal surgery, apical surgery, implant surgery, and surgical extractions of teeth (e.g., removal of erupted or nonerupted tooth requiring elevation of mucoperiosteal flap, removal of bone or section of tooth, and suturing if needed). During oral surgical procedures, dental practitioners should use only sterile solutions as a coolant or irrigant using an appropriate delivery device, such as a sterile bulb syringe, sterile tubing that bypasses dental unit waterlines, or sterile single-use devices.

### **Recommendations for Dental Providers**

- For all oral surgical procedures, use sterile saline or sterile water as a coolant or irrigant for surgical procedures.<sup>6</sup> Appropriate delivery devices can include:
  - bulb syringe,
  - sterile, single-use disposable products, or
  - sterile water delivery systems that bypass the dental unit by using sterile single-use disposable or sterilizable tubing.
- For all non-surgical dental procedures, use water that meets CDC recommendations (i.e.,  $\leq 500$  CFU/mL of heterotrophic water bacteria).

- For all non-surgical pulpal therapy and endodontic procedures, consider following more conservative recommendations from the following organizations which recommend irrigation with a sterile and/or antimicrobial solution:
  - the American Academy of Pediatric Dentistry (AAPD), [Pulp Therapy for Primary and Immature Permanent Teeth](#)
  - the American Association of Endodontists, [AAE Position Statement on Vital Pulp Therapy](#), and
  - the Organization for Safety, Asepsis and Prevention (OSAP), [Dental Unit Water Quality White Paper and Recommendations](#)
- Consult with the dental unit manufacturer for appropriate methods and equipment to maintain the quality of dental water. Many commercial devices and procedures are available and designed for this purpose.<sup>6</sup>
- Follow recommendations for monitoring water quality provided by the manufacturer of the unit or waterline treatment product.<sup>6</sup>
- Discharge water and air for a minimum of 20–30 seconds after each patient from any device connected to the dental water system that enters the patient’s mouth (e.g., handpieces, ultrasonic scalers, and air/water syringes).<sup>6</sup>
- Review the US Food and Drug Administration’s (FDA) website on [Dental Unit Waterlines](#) for recommendations for dental practitioners.

#### **Recommendations for Dental Facilities**

- Ensure that the dental facility has an infection prevention plan that includes policies and standard operating procedures dedicated to maintaining and monitoring water quality.<sup>8</sup>
- Provide staff training on how to properly maintain and monitor dental water quality. Training should be based on the manufacturer’s instructions for use of the products and devices used in the dental facility, provided for all new hires, and provided when new equipment is purchased and then at least annually.<sup>8</sup>
  - Contact the manufacturer of the treatment product or device if you have questions about the instructions for use.
- Document all maintenance records, monitoring results, and employee trainings. Accurate record keeping is an important component of a dental infection prevention program, ensures proper protocols have been met, and establishes accountability. Records should be maintained according to state and federal requirements.<sup>8</sup>
- Report infections suspected to be associated with receiving health care, including dental care, to the appropriate public health authorities. CDC provides [contact information for State Healthcare Associated Infections \(HAI\) Prevention Programs](#).

#### **Recommendations for Medical Providers**

- Consider an odontogenic source for cervical lymphadenitis or lymphadenopathy, and assess a history of dental care and procedures in all children presenting with lymphadenitis.

#### **Recommendations for Health Departments**

- During a healthcare-associated outbreak investigation of NTM infections of the head and neck region, assess patients for a history of recent dental treatment.
- Collaborate with [State Oral Health Programs](#) on investigations of healthcare-associated infections resulting from dental treatment.

#### **Recommendations for the Public**

- Contact your dental provider if you or your child develop an infection following dental treatment. Signs and symptoms of a postoperative dental infection could include a localized oral abscess, fever, or pain and swelling in the mouth or neck.
- Talk to your dental provider about their infection prevention and control practices and the steps their staff take to ensure safe treatment for all patients.

#### **For More Information**

- [Guidelines for Infection Control in Dental Healthcare Settings – 2003](#)
- [Dental Unit Waterlines \(FDA\)](#)
- [Policy on Infection Control \(American Academy of Pediatric Dentistry\)](#)
- [Dental Unit Water Quality: Organization for Safety, Asepsis and Prevention White Paper and Recommendations– 2018 \(OSAP\)](#)
- [Summary of Infection Prevention Practices in Dental Settings: Basic Expectations for Safe Care](#)

- [State-based HAI Prevention Activities](#)
- [Dental Unit Waterlines, Frequently Asked Questions](#)
- [Foundations: Building the Safest Dental Visit](#)

#### References

1. Singh J, O'Donnell K, Nieves DJ, et al. Invasive *Mycobacterium abscessus* Outbreak at a Pediatric Dental Clinic. *Open Forum Infect Dis*. 2021 Apr 15;8(6):ofab165. <https://pubmed.ncbi.nlm.nih.gov/34113683/>
2. Hatzenbuehler LA, Tobin-D'Angelo M, Drenzek C, et al. Pediatric Dental Clinic-Associated Outbreak of *Mycobacterium abscessus* Infection. *J Pediatric Infect Dis Soc*. 2017 Sep 1;6(3):e116-e122. <https://pubmed.ncbi.nlm.nih.gov/28903524/>
3. Pérez-Alfonzo R, Poleo Brito LE, Vergara MS, et al. Odontogenic cutaneous sinus tracts due to infection with nontuberculous mycobacteria: a report of three cases. *BMC Infect Dis*. 2020 Apr 21;20(1):295. <https://pubmed.ncbi.nlm.nih.gov/32316920/>
4. Bar T, Mishal J, Lewkowicz A, Nahlieli O. Osteomyelitis of the mandible due to *Mycobacterium abscessus*: a case report. *J Oral Maxillofac Surg*. 2005 Jun;63(6):841-4. <https://pubmed.ncbi.nlm.nih.gov/15944984/>
5. Ricci ML, Fontana S, Pinci F, et al. Pneumonia associated with a dental unit waterline. *Lancet* 2012; 379(9816):684. <https://pubmed.ncbi.nlm.nih.gov/22340301/>
6. Kohn WG, Collins AS, Cleveland JL, Harte JA, Eklund KJ, Malvitz DM; Centers for Disease Control and Prevention (CDC). Guidelines for infection control in dental health-care settings—2003. *MMWR Recomm Rep* 2003;52(RR-17):1-61. <https://www.cdc.gov/mmwr/PDF/rr/rr5217.pdf>
7. US Environmental Protection Agency. National Primary Drinking Water Regulations. [https://www.epa.gov/sites/default/files/2016-06/documents/npwdr\\_complete\\_table.pdf](https://www.epa.gov/sites/default/files/2016-06/documents/npwdr_complete_table.pdf)
8. Summary of Infection Prevention Practices in Dental Settings: Basic Expectations for Safe Care. Atlanta, GA: US Department of Health and Human Services, Centers for Disease Control and Prevention, National Center for Chronic Disease Prevention and Health Promotion, Division of Oral Health; March 2017. <https://www.cdc.gov/oralhealth/infectioncontrol/pdf/safe-care2.pdf>

*The Centers for Disease Control and Prevention (CDC) protects people's health and safety by preventing and controlling diseases and injuries; enhances health decisions by providing credible information on critical health issues; and promotes healthy living through strong partnerships with local, national and international organizations.*

DEPARTMENT OF HEALTH AND HUMAN SERVICES

HAN Message Types

- **Health Alert:** Conveys the highest level of importance; warrants immediate action or attention.
- **Health Advisory:** Provides important information for a specific incident or situation; may not require immediate action.
- **Health Update:** Provides updated information regarding an incident or situation; unlikely to require immediate action.
- **Info Service:** Provides general information that is not necessarily considered to be of an emergent nature.

###

This message was distributed to state and local health officers, state and local epidemiologists, state and local laboratory directors, public information officers, HAN coordinators, and clinician organizations.

###